



## OFFICE OF THE PRINCIPAL -CUM-CHIEF SUPERINTENDENT::GAUHATI MEDICAL COLLEGE HOSPITAL::GUWAHATI::32

No. GMC/MICRO/JE/2022 / 3 5 3

Dated: 3-3.2022

## CORRIGENDUM FOR WEBSITE PUBLICATION

## 1. Gel Documentation System

Previous specification				Modification done	
Magnetic	Handheld	washer	and	Magnetic Handheld washer and microplate shak	er
microplate shaker are included				are included are to be excluded	

## 2. Multiplex Bead based Cytokine Assay Platform

Previous specification  Validation kit should be provided with system for IQ/OQ and monthly validation of Optics, Fluidics, Classifier and Reporter channel for ensuring the instrument is running as per specification.  Blank polystyrene and magnetic beads alongwith amine coupling kit must be available with the vendor and should be quoted in optional for developing in-house assays. Please attach complete price list in USD and Indian Rupees for these beads and coupling kits.	be provided with system for IQ/OR for ensuring that the instrument is running as per specifications. Relevant technical notes/document must be attached stating the use and application of validation kit.
--	--



Should have x-plex platform for designing of pre-mixed custom panels.  Accessories like magnetic handheld washer and microplate shaker were not included	Should have x-plex/ procarta plex for designing of pre-mixed custom panels  Accessories like magnetic hand-held washer and microplate shaker are to be included
--	---

(Dr. A.C. Baishya)
Principal-cum-chief Superintendent
Gauhati Medical College
Guwahati – 32

Principal cum-chief Superintendant Gauhati Medicai Collego & Hospital & Guwanati-32